

Certificate

Self Test Mastercycler® pro Thermoblock

Mastercycler pro Model: MC pro S

Serial number: 6325XJ200131

Firmware software version: 3.003

fulfills requirements according to selftest procedures and manufacturer's thermoblock specifications:

Block Zone	Temperature Control (95 °C, 75 °C, 55 °C, 35 °C, 4 °C)	Ramping Rates
LEFT	PASSED	PASSED
CENTER	PASSED	PASSED
RIGHT	PASSED	PASSED

This certificate is valid until (month/day/year): _____

Nov/20/2008

User _____

Signature _____

Date



eppendorf
In touch with life

Your local distributor: www.eppendorf.com/worldwide
 Eppendorf AG Germany Tel: +49 40 538 01-0 - E-Mail: eppendorf@eppendorf.com
 Eppendorf North America, Inc. USA Tel: +1 516 334 7500 Toll free phone: +1 800 645 3050 E-Mail: info@eppendorf.com
 Application Support: E-Mail: support@eppendorf.com