

Service Check List



Eppendorf BioPhotometer® & Eppendorf BioSpectrometer® Essential Check

Instrument:	Instrument nick name:	Serial number:
Firmware version: Additional instrument software:	Installation date:	Location:
Job number:	Contract number:	Contact person:

CLEANING

Cleaning of instrument (housing and cuvette shaft)	yes		no	
--	-----	--	----	--

VISUAL INSPECTION

Checking for damages	---		passed		failed	
Inspection of display	---		passed		failed	
Inspection of housing and testing of stability	---		passed		failed	

FUNCTIONAL CHECK

Keyboard test	---		passed		failed	
Printer output test	---		passed		failed	
PC output test (for Backup and SW-Update via USB)	---		passed		failed	
dsDNA test method with 260 nm test filter	---		passed		failed	
Self test	n.a.		passed		failed	

COMMENTS

Completed service sticker attached to instrument case			YES		NO	

Date completed:	Service Engineer:	Service Engineer Signature:
	Customer:	Customer Signature:

Your local distributor: www.eppendorf.com/contact
 Eppendorf AG · 22331 Hamburg · Germany
 E-mail: eppendorf@eppendorf.com

Eppendorf®, the Eppendorf logo, Eppendorf BioPhotometer®, Eppendorf BioSpectrometer® and epServices® logo are registered trademarks of Eppendorf AG, Germany. U.S. Design Patents are listed on www.eppendorf.com/ip. All rights reserved incl. graphics and pictures. GB1/1215. Copyright 2015 © by Eppendorf AG.